VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100405 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:10/01/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.81: 0.079
LOT #: 33609080A1
EXPIRATION: 01/01/2012

BLANK TEST	0.000	20:51
INTERNAL STANDARD	VERIFIED	20:51
EXTERNAL STANDARD	0.079	20:51
BLANK TEST	0.000	20:52
EXTERNAL STANDARD	0.078	20:52
BLANK TEST	0.000	20:53
EXTERNAL STANDARD	0.078	20:53
BLANK TEST	0.000	20:54
EXTERNAL STANDARD	0.078	20:54
BLANK TEST	0.000	20:55
EXTERNAL STANDARD	0.079	20:55
BLANK TEST	0.000	20:56

Average = 0.0784
Std Dev = 0.0005

Diagnostic Check

VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.1°C	PASSED

PUMP INFO

Flow Rate = 4.795 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 11/4/11

Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 4 day of Nov, 20 11

Nikki Roth

Nikki Roth, Notary Public
My Commission Expires With Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

BMB 10/26/11