# VERIFICATION OF CALIBRATION REPORT

***of DataMaster dmt Breath Test Instrument***
***State of Alaska***

Serial #: 100412 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:07/18/2011

## External Standard Test Values

```
EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.90: 0.079
LOT #: 02511080A3
EXPIRATION: 02/01/2013

BLANK TEST                 0.000  12:02
INTERNAL STANDARD       VERIFIED  12:02
EXTERNAL STANDARD          0.077  12:02
BLANK TEST                 0.000  12:03
EXTERNAL STANDARD          0.077  12:03
BLANK TEST                 0.000  12:04
EXTERNAL STANDARD          0.077  12:04
BLANK TEST                 0.000  12:05
EXTERNAL STANDARD          0.077  12:05
BLANK TEST                 0.000  12:06
EXTERNAL STANDARD          0.077  12:06
BLANK TEST                 0.000  12:07

Average = 0.0770
Std Dev = 0.0000
```

## Diagnostic Check

```
VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C          PASSED
Breath Tube    = 47.0°C          PASSED

PUMP INFO
Flow Rate = 5.053 L/M            PASSED

DETECTOR INFO
PUMP ON                          PASSED
PUMP OFF                         PASSED

FILTER INFO
Filter 1                         PASSED
Filter 2                         PASSED
Filter 3                         PASSED

INTERNAL STANDARD                PASSED
```

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 8/1/11
Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1 day of Aug, 20 11

Nikki Roth — Nikki Roth
Notary Public's Signature — Notary Name
My Commission Expires on With Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

BMB 7/22/11