

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130179</u> ✓	
A	Name <u>John T. Ostoj</u> ID# <u>4823</u> Date <u>11-24-08</u> ✓
	Agency <u>Soldotna AST</u> Phone # <u>262-4453</u>
	Instrument Location <u>Soldotna AST Post</u>
B	Alco S/N <u>X301556</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>600</u> ✓
	<u>.078</u> Adjusted.
	Alco Test Value Average <u>.078</u> ✓
	1 st Alco <u>.078</u> ✓
	2 nd Alco <u>.078</u> ✓
	Signature <u>J. T. Ostoj</u> ✓
	<u>CSB</u> <u>12/5/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, John Ostoj, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

12/5/08
Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

NOVEMBER 24, 2008 ✓ *COB 12/5/08*

OPERATOR'S NAME:

DSTOJ/JOHN/T

OPERATOR'S NUMBER: 4823

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: UOC TEST

DEPT/AGENCY: SLD00

CASE/REPORT: UOC TEST

TEST TYPE: U

ALCO TARGET VALUE: .081 ✓

ALCO S/N: 8301306

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.12 in		
ALCO TARGET	.078	07:43
BLANK TEST	.000	07:44
INTERNAL STANDARD	VERIFIED	07:44
ALCO TU 29.12 in	.078 ✓	07:45
BLANK TEST	.000	07:46
SUBJECT SAMPLE <i>COB 12/5/08</i>	.000	07:46
BLANK TEST	.000	07:47
ALCO TU 29.12 in	.078 ✓	07:47
BLANK TEST	.000	07:48

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

NOVEMBER 24, 2008 ✓
TIME 07:41

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.12 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()*+,-./:;<=>?@BCDEFG
HIJKL MNOPQRS TUVWXYZ 1234567890
pqrs tuvwxyz 1234