

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | | | |
|--|------------------|--|--|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N | <u>130182</u> ✓ |
| Name | | <u>Tim Tuckwood</u> | ID# <u>3446</u> Date <u>12/16th 08</u> ✓ |
| A | Agency | <u>Alaska State Troopers</u> | Phone # _____ |
| Instrument Location | | <u>Delta Courthouse</u> | ✓ |
| Alco S/N | | <u>X301596</u> ✓ | Target Value <u>0.081</u> ✓ High Pressure <u>750 psi</u> |
| B | Alco Test Values | <u>0.080</u> ✓ 1 st Alco | <u>0.080</u> ✓ 2 nd Alco |
| Signature | | <u><i>Tim Tuckwood</i></u> ✓ | <u>003</u> <u>1/6/09</u> |

(OVER)

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Tim Tuckwood, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

1/6/09
Date

JAN 02 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

DECEMBER 17, 2008 *CSB 1/10/09*

OPERATOR'S NAME:
TUCKWOOD TIM
OPERATOR'S NUMBER: 3446
SUBJECT'S LAST NAME:
VERIF
SUBJECT'S FIRST NAME/MI :
OF CAL
O.L. #: 12345
DEPT/AGENCY: DJTO
CASE/REPORT: 09876
TEST TYPE: U
ALCO TARGET VALUE: .081 ✓
ALCO S/N: X301596

--- BREATH ANALYSIS ---

| | | |
|-----------------------------------|----------|-------|
| .081 ADJUSTED FOR 29.42 in | | |
| ALCO TARGET | .079 | 10:56 |
| BLANK TEST | .000 | 10:57 |
| INTERNAL STANDARD | VERIFIED | 10:57 |
| ALCO TV 29.42 in | .080 ✓ | 10:58 |
| BLANK TEST | .000 | 10:58 |
| SUBJECT SAMPLE <i>CSB 1/10/09</i> | .080 | 10:59 |
| BLANK TEST | .000 ✓ | 10:59 |
| ALCO TV 29.42 in | .080 | 11:00 |
| BLANK TEST | .000 | 11:01 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130182

DECEMBER 17, 2008 ✓
TIME 10:53

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 41c |
| BAROMETER: | 29.42 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHI
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~■