

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
 State of Alaska
 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 01 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130183</u> ✓	7/12/10 BMB
Name		<u>HUTT, William T.</u>	ID#	<u>3359</u>
		Date	<u>06/30/09</u>	10
A	Agency	<u>Homer Police Dept./Jail</u>		Phone #
				<u>235-3150</u>
	Instrument Location	<u>4060 Heath St. Homer, AK</u>		Fax
				<u>235-3151</u>
				<u>99603</u>
B	Alco S/N	<u>A090740</u> ✓	Target Value	<u>0.084</u> ✓
			High Pressure	<u>750 psi</u>
	Alco Test Values	<u>.083</u> ✓	<u>.084</u> ✓	
		1 st Alco	2 nd Alco	
	Signature	<u>William T. Hutt</u> ✓		

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 21st day of July, 2010.

Carolyn M. Noland
 Carolyn M. Noland
 Notary Public, State of Alaska
 Commission Expires with Office



BMB
7/12/10

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130183
 JUNE 30, 2010 ✓ *bmb 7/12/10*
 OPERATOR'S NAME: HOTT, WILLIAM T
 OPERATOR'S NUMBER: 3359
 SUBJECT'S LAST NAME: VERIFICATION OF CAL
 SUBJECT'S FIRST NAME/NI: ETON
 O.L. #: 01234567
 DEPT/AGENCY: HOM1
 CASE/REPORT: 10-0000
 TEST TYPE: U
 ALCO TARGET VALUE: 0.084 ✓
 ALCO S/N: A090740 ✓

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.81 in		
ALCO TARGET	.083	02:56
BLANK TEST	.000	02:57
INTERNAL STANDARD	VERIFIED	02:57
ALCO TV 29.81 in	.083	02:58
BLANK TEST	.000	02:58
SUBJECT SAMPLE	.084	03:00
BLANK TEST	.000	03:00
ALCO TV 29.81 in	.084	03:01
BLANK TEST	.000	03:02

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
 DATAMASTER cdm 130183 ✓
 JUNE 30, 2010 ✓ *WTH*
 TIME 02:49

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.81 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
 H I J K L M N O P Q R S T U V W X Y Z I ^ _ ` a b c d e f g h i j k l m n o
 p q r s t u v w x y z () =