

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130190</u> ✓
A	Name	<u>STEVE DUNN</u>	ID# <u>4106</u> Date <u>4/14/08</u> ✓
	Agency	<u>APD</u>	Phone # <u>786-8900</u>
	Instrument Location	<u>JAIL</u>	
B	Alco S/N	<u>X124625</u> ✓	Target Value <u>.086</u> ✓ High Pressure <u>300</u>
	Alco Test Value Average	<u>.083</u> ✓ 1 st Alco	<u>.083</u> ✓ 2 nd Alco
Signature		 <u>COB</u> <u>4/14/08</u>	
(OVER)			

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Steve P Dunn, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/27/08
Date

APR 17 2008

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130190

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

S. DUNN #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

APRIL 14, 2008 *CDP #21108*

OPERATOR'S NAME:
DUNN/S
OPERATOR'S NUMBER: 4106
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
U
D.L. #: U
DEPT/AGENCY: ANCI
CASE/REPORT: U
TEST TYPE: U
ALCO TARGET VALUE: .086
ALCO S/N: X124625

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 29.83 in		
ALCO TARGET	.085	11:17
BLANK TEST	.000	11:18
INTERNAL STANDARD	VERIFIED	11:18
ALCO TU 29.83 in	.083	11:18
BLANK TEST	.000	11:19
SUBJECT SAMPLE	.000 <i>CDP #21108</i>	11:20
BLANK TEST	.000	11:20
ALCO TU 29.83 in	.083	11:21
BLANK TEST	.000	11:22

DIAGNOSTIC CHECK

S. DUNN #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130190

APRIL 14, 2008
TIME 11:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.83 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde{ghijklmno
pqrstuvwxyz{|}~