

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130306</u>
Name		<u>Derek Sitz</u>	ID# <u>4096</u> Date <u>11.6.08</u>
A	Agency	<u>Anchorage Police Department</u>	Phone # <u>786.8900</u>
Instrument Location		<u>Seely Substation</u>	
Alco S/N		<u>77198</u>	Target Value <u>.079</u> High Pressure <u>1200</u>
B	Alco Test Values	<u>.076</u> 1 st Alco	<u>.076</u> 2 nd Alco
Signature		<u>Derek Sitz</u>	<u>CSB</u> <u>11/20/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Derek Sitz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz 11/21/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 13030C ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

NOVEMBER 06, 2008

OPERATOR'S NAME:
SITZ, DEREK W
OPERATOR'S NUMBER: 4096
SUBJECT'S LAST NAME:
UOC

SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: 1234567
DEPT/AGENCY: ANCI
CASE/REPORT: 00-1234
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: 77198

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.78 in
ALCO TARGET .078 06:06
BLANK TEST .000 06:07
INTERNAL STANDARD VERIFIED 06:07
ALCO TV 29.78 in .076 ✓ 06:08
BLANK TEST .000 06:09
SUBJECT SAMPLE *DS* .000 06:09
BLANK TEST *11.6.08* .000 ✓ 06:10
ALCO TV 29.80 in .076 06:11
BLANK TEST .000 06:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306 ✓

NOVEMBER 06, 2008
TIME 06:13

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.80 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLWOPQRSTUUVWXYZ[\]^_`abcde fghijklmno
pqrstuwxvzy{|}~#