VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130308

Supervisor/Operator Performing the Verification Procedure:

Name Malik Jones ID# 5830 Date 7/12/08

A

Agency AST Phone # 451-5100

Instrument Location Healy

B

Alco S/N X173355 Target Value .082 High Pressure 650

Alco Test Value Average .078 (1st Alco) .077 (2nd Alco)

Signature [signature] 7/22/08

(OVER)

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Malik Jones, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

[signature] 7/22/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 03/00 JUL 18 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N - 130308

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130308

JULY 12, 2008 COB 7/22/08

OPERATOR'S NAME:
JONES/MALIK/JAMAL
OPERATOR'S NUMBER: 5830
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
0
O.L. #: 0
DEPT/AGENCY: HLY0
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: X173355

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 28.77 in

ALCO TARGET	.078	14:02
BLANK TEST	.000	14:03
INTERNAL STANDARD	VERIFIED	14:03
ALCO TV 28.77 in	.078	14:04
BLANK TEST	.000	14:04
SUBJECT SAMPLE	.000 COB 7/22/08	14:08
BLANK TEST	.000	14:09
ALCO TV 28.77 in	.077	14:09
BLANK TEST	.000	14:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130308

JULY 12, 2008
TIME 14:11

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	28.77 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~