

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130309
Name	PATRICK NELSON	ID#	4389 Date 5/19/08 ✓
A Agency	AST	Phone #	373-8337
Instrument Location	MATSU WEST, WASILLA ✓		
B Alco S/N	X172971 ✓	Target Value	.079 ✓ High Pressure 500
Alco Test Values	.079 ✓ 1 st Alco	.078 ✓ 2 nd Alco	
Signature	P Nelson		PSNO CWB 6/3/08

(OVER)

Under the Alaska Rules of Evidence, I certify that (in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



J Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/5/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

MAY 19, 2008

OPERATOR'S NAME:
NELSON PATRICK
OPERATOR'S NUMBER: 4389
SUBJECT'S LAST NAME:
VERIFICATION OF
SUBJECT'S FIRST NAME/MI :
CALIBRATION
O.L. #: 123456
DEPT/AGENCY: PL#0
CASE/REPORT: 12-34567
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: X172971

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.80 in
ALCO TARGET .078 19:37
BLANK TEST .000 19:38
INTERNAL STANDARD VERIFIED 19:38
ALCO TV 29.80 in .079 19:39
BLANK TEST .000 19:39
SUBJECT SAMPLE .000 19:40
BLANK TEST .000 19:41
ALCO TV 29.80 in .078 19:41
BLANK TEST .000 19:42

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

MAY 19, 2008
TIME 19:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.80 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■