

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130570 ✓		
Name	CHRIS RITALA	ID#	3791	Date	8/28/08 ✓
Agency	ANCHORAGE P.D.	Phone #	786.2640		
Instrument Location	FLORA SUB. STATION (4 <sup>TH</sup> AVE) ✓				
Alco S/N	58428 ✓	Target Value	.078 ✓	High Pressure	1050
Alco Test Values	.078 ✓		.079 ✓		
	1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco		
Signature			RITALA/1349 ✓		COB 9/2/08
(OVER)					

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

9/15/08

Date

SEP 03 2008

BT9.03.00

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

CDL  
9-11-00

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130570

AUGUST 28, 2000 ✓  
CDL 9/12/00

OPERATOR'S NAME:  
RITALA  
OPERATOR'S NUMBER: 3791  
SUBJECT'S LAST NAME:  
VERIFICATION OF CALIBRATION  
SUBJECT'S FIRST NAME/MI:  
A  
O.L. #: A  
DEPT/AGENCY: ANCI  
CASE/REPORT: A  
TEST TYPE: V  
ALCO TARGET VALUE: .078 ✓  
ALCO S/N: 58428 ✓

### --- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.85 in  
ALCO TARGET .077 17:52  
BLANK TEST .000 17:53  
INTERNAL STANDARD VERIFIED 17:53  
ALCO TV 29.85 in .078 ✓ 17:54  
BLANK TEST .000 17:54  
SUBJECT SAMPLE .000 ✓ 17:55  
CDL 9/12/00  
BLANK TEST .000 17:56  
ALCO TV 29.85 in .079 ✓ 17:56  
BLANK TEST .000 17:57

CR/1349 8/29/00

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130570

AUGUST 28, 2000 ✓  
TIME 17:48

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 51c  
BREATH TUBE: 48c  
BAROMETER: 29.85 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

! "#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF0  
HIJKLNPQRSTUWXYZ1234567890:;  
pqrstuvwxyz()

CR/1349 8/29/00