

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130571 ✓
Name		Tim Tuckwood ✓	ID# 3446 Date 4-9-08
A	Agency	Alaska State Troopers ✓	Phone # 895-4800
Instrument Location		Arctic Man / Summit Lake	
Alco S/N		31085 ✓	Target Value .084 ✓ High Pressure 1000 psi
B	Alco Test Values	.074 ✓ 1 st Alco	.074 ✓ 2 nd Alco
Signature		<i>Tim Tuckwood</i>	COS 4/15/08

(OVER)

Under the Alaska Rules of Evidence, I certify that (in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Tim Tuckwood, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/20/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130571

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

N

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

APRIL 09, 2008

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF

SUBJECT'S FIRST NAME/MI :

OF CA

O.L. #: 12345

DEPT/AGENCY: DJT0

CASE/REPORT: 09876

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: 31085

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 26.11 in

ALCO TARGET	.073	16:18
BLANK TEST	.000	16:19
INTERNAL STANDARD	VERIFIED	16:19
ALCO TV 26.11 in	.074	16:20
BLANK TEST	.000	16:20
SUBJECT SAMPLE	.000	16:21
BLANK TEST	.000	16:22
ALCO TV 26.11 in	.074	16:22
BLANK TEST	.000	16:23

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130571

APRIL 09, 2008

TIME 16:26

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	26.11 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z () * +