# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*
*State of Alaska*
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130579

Supervisor/Operator Performing the Verification Procedure:

Name C. RITALA /1349 ID# 3791 Date 7/19/08

A

Agency ANCHORAGE POLICE DEPT. Phone # 786-2640

Instrument Location ANCHORAGE JAIL #2

Alco S/N 57447 Target Value 071 High Pressure 1000

B

Alco Test Values 077 (1st Alco) 078 (2nd Alco)

Signature RITALA/1349 CDB 7/25/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

7/28/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9.03/00

JUL 21 2008

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*
*State of Alaska*
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

**DataMaster cdm S/N** 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JULY 19, 2008 CDB 7/26/08

OPERATOR'S NAME:
RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: ANC1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .071
ALCO S/N: 57447

--- BREATH ANALYSIS ---

.071 ADJUSTED FOR 29.70 in

| | | |
|---|---|---|
| ALCO TARGET | .070 | 23:59 |
| BLANK TEST | .000 | 00:00 |
| INTERNAL STANDARD | VERIFIED | 00:00 |
| ALCO TV 29.70 in | .077 | 00:01 |
| BLANK TEST | .000 | 00:02 |
| SUBJECT SAMPLE | .000 CDB 7/26/08 | 00:03 |
| BLANK TEST | .000 | 00:04 |
| ALCO TV 29.70 in | .078 | 00:04 |
| BLANK TEST | .000 | 00:05 |

30/85/1

CR/1349 7/19/08

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JULY 19, 2008
TIME 23:58

--- DIAGNOSTIC CHECK ---

| | |
|---|---|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| BREATH TUBE: | 42c |
| BAROMETER: | 29.70 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgh ijklmno
pqrstuvwxyz{|}~

CR/1349 7/19/08

BT9 03/0