

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100348

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/25/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.69: 0.079
LOT #: 02511080A3
EXPIRATION: 02/01/2013

BLANK TEST	0.000	23:39
INTERNAL STANDARD	VERIFIED	23:39
EXTERNAL STANDARD	0.080	23:39
BLANK TEST	0.000	23:40
EXTERNAL STANDARD	0.079	23:40
BLANK TEST	0.000	23:41
EXTERNAL STANDARD	0.080	23:42
BLANK TEST	0.000	23:42
EXTERNAL STANDARD	0.080	23:43
BLANK TEST	0.000	23:44
EXTERNAL STANDARD	0.080	23:44
BLANK TEST	0.000	23:45

Average = 0.0798
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 44.8°C PASSED

PUMP INFO

Flow Rate = 5.167 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

4/15/11

Subscribed and sworn before me this 15 day of JUNE, 20 11





Notary Public's Signature
My Commission Expires on WITH OFFICE

KYLE KLOPFER
Notary Name


4/29/11