

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/14/2011

Serial #: 100352

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.25: 0.078  
LOT #: 33609080A1  
EXPIRATION: 01/01/2012

BLANK TEST	0.000	08:15
INTERNAL STANDARD	VERIFIED	08:15
EXTERNAL STANDARD	0.077	08:15
BLANK TEST	0.000	08:16
EXTERNAL STANDARD	0.077	08:17
BLANK TEST	0.000	08:17
EXTERNAL STANDARD	0.077	08:18
BLANK TEST	0.000	08:19
EXTERNAL STANDARD	0.077	08:19
BLANK TEST	0.000	08:20
EXTERNAL STANDARD	0.077	08:20
BLANK TEST	0.000	08:21

Average = 0.0770  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 1.01  
PIC: 3.02  
Modem: 2.0  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.3°C	PASSED

### PUMP INFO

Flow Rate = 5.035 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.




Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

5/13/11

Subscribed and sworn before me this 13 day of MAY, 20 11

  
Notary Public's Signature  
My Commission Expires on WITH

KYLE KLOPPER  
Notary Name  
OFFICE



003/14/11