

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100354

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/09/2012

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.03: 0.077
LOT #: 07511080A1
EXPIRATION: 03/01/2013

BLANK TEST	0.000	11:36
INTERNAL STANDARD	VERIFIED	11:36
EXTERNAL STANDARD	0.076	11:37
BLANK TEST	0.000	11:37
EXTERNAL STANDARD	0.075	11:38
BLANK TEST	0.000	11:39
EXTERNAL STANDARD	0.075	11:39
BLANK TEST	0.000	11:40
EXTERNAL STANDARD	0.075	11:40
BLANK TEST	0.000	11:41
EXTERNAL STANDARD	0.075	11:41
BLANK TEST	0.000	11:42

Average = 0.0752
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 2.00
PIC: 3.03
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED
Breath Tube = 40.5°C PASSED

PUMP INFO

Flow Rate = 3.995 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

4/6/2012

Subscribed and sworn before me this 6 day of April, 20 12


Nikki Roth, Notary Public
My Commission Expires With Office



BMB
3/12/12