# VERIFICATION OF CALIBRATION REPORT

***of DataMaster dmt Breath Test Instrument***
***State of Alaska***

Serial #: 100358 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:03/29/2012

## External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.05: 0.077
LOT #: 02511080A2
EXPIRATION: 02/01/2013

| | | |
|---|---|---|
| BLANK TEST | 0.000 | 15:40 |
| INTERNAL STANDARD | VERIFIED | 15:40 |
| EXTERNAL STANDARD | 0.076 | 15:40 |
| BLANK TEST | 0.000 | 15:41 |
| EXTERNAL STANDARD | 0.076 | 15:41 |
| BLANK TEST | 0.000 | 15:42 |
| EXTERNAL STANDARD | 0.076 | 15:42 |
| BLANK TEST | 0.000 | 15:43 |
| EXTERNAL STANDARD | 0.077 | 15:44 |
| BLANK TEST | 0.000 | 15:44 |
| EXTERNAL STANDARD | 0.077 | 15:45 |
| BLANK TEST | 0.000 | 15:46 |

Average = 0.0764
Std Dev = 0.0005

## Diagnostic Check

VERSIONS
DMT: 2.00
PIC: 3.03
Modem: 2.3
Questions: 2.2

TEMPERATURES

| | |
|---|---|
| Sample Chamber = 49.1°C | PASSED |
| Breath Tube = 41.1°C | PASSED |

PUMP INFO

| | |
|---|---|
| Flow Rate = 4.909 L/M | PASSED |

DETECTOR INFO

| | |
|---|---|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|---|---|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

| | |
|---|---|
| INTERNAL STANDARD | PASSED |

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 5/21/2012
Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of May, 20 12

Nikki Roth
**Nikki Roth, Notary Public**
**My Commission Expires With Office**

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

BMB 5/1/12