VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100364 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:08/03/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.69: 0.079
LOT #: 07511080A1
EXPIRATION: 03/01/2013

BLANK TEST	0.000	13:06
INTERNAL STANDARD	VERIFIED	13:06
EXTERNAL STANDARD	0.077	13:07
BLANK TEST	0.000	13:08
EXTERNAL STANDARD	0.078	13:08
BLANK TEST	0.000	13:09
EXTERNAL STANDARD	0.078	13:09
BLANK TEST	0.000	13:10
EXTERNAL STANDARD	0.077	13:10
BLANK TEST	0.000	13:11
EXTERNAL STANDARD	0.078	13:11
BLANK TEST	0.000	13:12

Average = 0.0776
Std Dev = 0.0005

Diagnostic Check

VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 42.7°C	PASSED

PUMP INFO

Flow Rate = 5.277 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J Bolz 8/12/11
Nita J Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug. 20 11

Nikki Roth — Nikki Roth
Notary Public's Signature — Notary Name
My Commission Expires on With Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BMB 8/8/11