VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Serial #: 100365 Date:03/25/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.49: 0.078
LOT #: 33609080A1
EXPIRATION: 01/01/2012

BLANK TEST	0.000	15:19
INTERNAL STANDARD	VERIFIED	15:19
EXTERNAL STANDARD	0.079	15:19
BLANK TEST	0.000	15:20
EXTERNAL STANDARD	0.079	15:21
BLANK TEST	0.000	15:22
EXTERNAL STANDARD	0.078	15:22
BLANK TEST	0.000	15:23
EXTERNAL STANDARD	0.079	15:23
BLANK TEST	0.000	15:24
EXTERNAL STANDARD	0.079	15:24
BLANK TEST	0.000	15:25

Average = 0.0788
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 43.9°C	PASSED

PUMP INFO

Flow Rate = 4.558 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

6/15/11

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 20 11

Notary Public's Signature
My Commission Expires: WITH OFFICE

KYLE KLOPFER
Notary Name

KYLE KLOPFER
NOTARY PUBLIC
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

003 3/30/11