VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100369 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:03/18/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.19: 0.080
LOT #: 02511080A1
EXPIRATION: 02/01/2013

BLANK TEST	0.000	11:55
INTERNAL STANDARD	VERIFIED	11:55
EXTERNAL STANDARD	0.080	11:55
BLANK TEST	0.000	11:56
EXTERNAL STANDARD	0.080	11:56
BLANK TEST	0.000	11:57
EXTERNAL STANDARD	0.080	11:57
BLANK TEST	0.000	11:58
EXTERNAL STANDARD	0.079	11:58
BLANK TEST	0.000	11:59
EXTERNAL STANDARD	0.080	12:00
BLANK TEST	0.000	12:01

Average = 0.0798
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 42.9°C	PASSED

PUMP INFO

Flow Rate = 6.010 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

5/13/11

Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 13 day of MAY, 20 11

Notary Public's Signature
My Commission Expires on WITH OFFICE

KYLE KLOPFER
Notary Name

KYLE KLOPFER
NOTARY PUBLIC
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

3/23/11