

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100382

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/18/2011

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.98: 0.080  
LOT #: 33609080A1  
EXPIRATION: 01/01/2012

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | 0.000    | 14:02 |
| INTERNAL STANDARD | VERIFIED | 14:03 |
| EXTERNAL STANDARD | 0.080    | 14:03 |
| BLANK TEST        | 0.000    | 14:04 |
| EXTERNAL STANDARD | 0.080    | 14:04 |
| BLANK TEST        | 0.000    | 14:05 |
| EXTERNAL STANDARD | 0.080    | 14:05 |
| BLANK TEST        | 0.000    | 14:06 |
| EXTERNAL STANDARD | 0.080    | 14:06 |
| BLANK TEST        | 0.000    | 14:07 |
| EXTERNAL STANDARD | 0.080    | 14:08 |
| BLANK TEST        | 0.000    | 14:08 |

Average = 0.0800  
Std Dev = 0.0000

## Diagnostic Check

VERSIONS  
DMT: 1.01  
PIC: 3.02  
Modem: 2.0  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C PASSED  
Breath Tube = 44.8°C PASSED

PUMP INFO  
Flow Rate = 5.274 L/M PASSED

DETECTOR INFO  
PUMP ON PASSED  
PUMP OFF PASSED

FILTER INFO  
Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

6/15/11

Subscribed and sworn before me this 15 day of JUNE, 20 11

  
Notary Public's Signature  
My Commission Expires on WITH OFFICE

Kyle Klopfer  
Notary Name



008/03/11