VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100384 — *Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program* — Date:09/06/2012

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.37: 0.081
LOT #: 07511080A1
EXPIRATION: 03/01/2013

BLANK TEST	0.000	10:40
INTERNAL STANDARD	VERIFIED	10:40
EXTERNAL STANDARD	0.079	10:40
BLANK TEST	0.000	10:41
EXTERNAL STANDARD	0.079	10:41
BLANK TEST	0.000	10:42
EXTERNAL STANDARD	0.079	10:42
BLANK TEST	0.000	10:43
EXTERNAL STANDARD	0.079	10:43
BLANK TEST	0.000	10:44
EXTERNAL STANDARD	0.079	10:44
BLANK TEST	0.000	10:45

Average = 0.0790
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 2.00
PIC: 3.02
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 47.5°C	PASSED

PUMP INFO

Flow Rate = 4.799 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 10/4/2012

Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 4th day of Oct, 20 12

Nikki Roth

Nikki Roth, Notary Public
My Commission Expires With Office

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

COB 9/14/12