# VERIFICATION OF CALIBRATION REPORT

***of DataMaster dmt Breath Test Instrument***
***State of Alaska***

Serial #: 100421 — ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** — Date:12/12/2011

## External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.32: 0.078
LOT #: 07511080A1
EXPIRATION: 03/01/2013

| | | |
|---|---|---|
| BLANK TEST | 0.000 | 12:54 |
| INTERNAL STANDARD | VERIFIED | 12:54 |
| EXTERNAL STANDARD | 0.077 | 12:54 |
| BLANK TEST | 0.000 | 12:55 |
| EXTERNAL STANDARD | 0.078 | 12:55 |
| BLANK TEST | 0.000 | 12:56 |
| EXTERNAL STANDARD | 0.078 | 12:56 |
| BLANK TEST | 0.000 | 12:57 |
| EXTERNAL STANDARD | 0.078 | 12:57 |
| BLANK TEST | 0.000 | 12:58 |
| EXTERNAL STANDARD | 0.079 | 12:58 |
| BLANK TEST | 0.000 | 12:59 |

Average = 0.0780
Std Dev = 0.0007

## Diagnostic Check

VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

| | |
|---|---|
| Sample Chamber = 48.8°C | PASSED |
| Breath Tube = 43.6°C | PASSED |

PUMP INFO

| | |
|---|---|
| Flow Rate = 4.858 L/M | PASSED |

DETECTOR INFO

| | |
|---|---|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|---|---|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

| | |
|---|---|
| INTERNAL STANDARD | PASSED |

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 1/25/12

Nita J.Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 25 day of Jan 20 12

Nikki Roth

**Nikki Roth, Notary Public**
**My Commission Expires With Office**

NIKKI ROTH
My Commission Expires
NOTARY PUBLIC
With Office
STATE OF ALASKA

STATE OF ALASKA
SCIENTIFIC CRIME DETECTION LABORATORY

BMB 12/21/11