

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

DOCUMENT STRUCTURE

SECTION 1: INTRODUCTION	2
SECTION 2: EVIDENCE HANDLING	3
SECTION 3: ACCEPTANCE REQUIREMENTS.....	5
SECTION 4: DILUENT, CALIBRATION STANDARDS AND CONTROLS.....	6
SECTION 5: EQUIPMENT AND INSTRUMENTATION.....	10
SECTION 6: ANALYSIS PROCEDURE	11
SECTION 7: REPORTING	14
SECTION 8: CASE FILES.....	16
SECTION 9: PROFICIENCY TESTING.....	17
SECTION 10: PROVIDING OPINIONS FOR COURT PROCEEDINGS.....	18
SECTION 11: TOXICOLOGY SEND OUTS	19
SECTION 12: ABBREVIATIONS.....	20
APPENDIX A: UNCERTAINTY OF MEASUREMENT	21
APPENDIX B: REVISION HISTORY	22
APPENDIX C: SAMPLING PLAN.....	23

ARCHIVED 7/19/2015

SECTION 1: INTRODUCTION

The Alaska Administrative Code dictates aspects of the collection of blood and method of alcohol analysis. 13 AAC 063.110 gives information on the collection and handling of blood samples and 13 AAC 063.120 details methods of blood alcohol analysis that are appropriate for use in the State of Alaska.

The method used by this laboratory for quantitative ethanol analysis of body fluids and beverages is headspace gas chromatography with a flame ionization detector and mass spectrometer. A computer-interfaced instrument calibrates and analyzes specimens within a programmed sequence. The result is a comparison of unknown case samples with known concentrations of ethanol, with any variations in amount of injection corrected by the addition of an internal standard, n-propanol, in the diluent.

ARCHIVED 7/9/2015

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

SECTION 2: EVIDENCE HANDLING

Evidence submitted to the Laboratory for ethanol analysis includes but is not limited to blood, beverage samples and homebrews. Blood is usually submitted in specific forensic alcohol collection kits containing gray top Vacutainer® tubes. Gray top tubes contain a sodium fluoride preservative and potassium oxalate anticoagulant. Gray top tubes do not require refrigeration because of their preservatives, but collection kits are refrigerated upon receipt to provide extra stability for any unpreserved samples and to comply with the Alaska Administrative Code. Forensic alcohol collection kits are also acceptable for submitting beverage and homebrew samples. Beverage samples are not required to be refrigerated.

Evidence is stored in a locked refrigerator at all times except during analysis. During analysis the evidence must be secured in the locked blood alcohol laboratory if the analyst is not present.

When opening evidence for analysis each layer of sealed packaging should be marked with the laboratory number, item number, date it was opened and the initials of the analyst who opened it. The tube or item that is to be analyzed should be marked with the laboratory number, item number, date opened, and initials.

After opening the packaging, each blood sample should be checked for a subject name which should in turn be checked against the Request for Laboratory Services form for accuracy. Any discrepancies or lack of name will be documented in the analyst's notes.

The analyst's notes should also include: how the sample was packaged, if the packaging was sealed, the number and type of tubes submitted, any unusual aspects of the sample, and any broken seals or discrepancies. The date and time of collection and subject name should be recorded if that information is available from the blood tube or sealed container. The analyst should also record the date that the evidence is opened as the start of analysis date in the appropriate field of the notes function in the LIMS. The date of completion field should be left blank until the analyst has completed their analysis and is ready to forward the report on for technical review. The date of completion is entered when the analyst is ready to roll the draft complete milestone in the LIMS.

Outer seals must be resealed after analysis with tamper-proof evidence tape. There should be some means of identifying when and by whom this was performed, usually by initialing and dating the evidence tape. Samples that require toxicology testing should be packaged for shipment to Washington State Patrol Toxicology Laboratory. The outer packaging and any additional tubes should be resealed and returned to the evidence section along with all samples that do not require toxicology testing.

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

Note: Blood and urine are biological materials and precautions associated with handling a biological hazard should be taken. Please refer to the Laboratory's Health and Safety Manual for more information regarding the handling of biological materials.

ARCHIVED 7/19/2015

SECTION 3: ACCEPTANCE REQUIREMENTS

Acceptability Requirements

The aqueous controls, whole blood controls, beverage controls and case samples are run in duplicate and calculated in the same manner. Duplicate results are truncated to 4 decimal places and averaged. The difference between the truncated 4 decimal place numbers is divided by the average to determine the RPD. The RPD must be 3% or less for the average to be used.

For aqueous and whole blood controls the average truncated to 3 decimal places is compared to the expected value. The passing criteria are equal to the expanded coverage probability of +/-4.2% of the expected value converted to g/100 mL units rounded to three decimal places. The beverage control value is converted to alcohol by volume as outlined in Section 6, Sample preparation, and the value must fall within 4.3% of the expected value.

Ethanol Control Worksheet

This is a separately controlled document which contains control lot information as well as passing criteria.

The Ethanol Control Worksheet is placed in the LIMS as part of the control pack for each batch of samples analyzed. The Ethanol Control Worksheet is also used to track all control data for future uncertainty evaluations. The values and lot numbers for all calibrators and controls as well as ID and current use for the auto dilutors are listed on the Ethanol Control Worksheet. This must be entered into the LIMS where they will become part of each case file as a permanent record of the control used for each batch analysis.

ARCHIVED 7/19/2015

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

SECTION 4: DILUENT, CALIBRATION STANDARDS AND CONTROLS

Internal Standard Diluent

Chemicals that meet American Chemical Society (A.C.S.) specifications will be used when available. Mix 0.2 mL of n-propanol in 2 liters of deionized water. Solution can be stored at room temperature and expires one year from date of preparation.

A newly prepared internal standard diluent is deemed suitable for use in casework when no ethanol is detected in the negative control and the mass spectrum identifies n-propanol when compared with a reference spectrum. The internal standard diluent's preparation is documented in the blood alcohol discipline's reagent log.

After a new lot of internal standard diluent is verified the following documentation is included in the blood alcohol discipline's reagent log:

1. The chromatogram of the negative control showing n-propanol and no ethanol is detected
2. The mass spectrum of n-propanol with a library match

Calibration Standards

NIST traceable aqueous ethanol standards at 0.020, 0.100, 0.200 and 0.500 g/100 mL are used for calibration. Calibration standards are purchased from an ISO/IEC 17025:2005 certified supplier. A critical supply vendor form is completed for each calibration standard. The Certificates of Analysis for all calibration standards are kept in the blood alcohol laboratory.

Prior to using a new lot of calibration standard one vial from the lot should be run as a sample (in duplicate) to verify the lot falls within the manufacturer's stated concentration +/- the expanded coverage probability of 4.2%. See Section 3 Acceptability Requirements for further details. A new shipment of the same lot does not require additional testing. The lot numbers and expiration dates should then be entered into the LIMS where they will become part of each case file as a permanent record of the standards used for each calibration.

After a new lot of calibration standard is verified, the following documentation is included in the Calibration Standards binder:

1. Certificate of Analysis
2. Control Pack (See Section 8 for requirements)
3. ETOH custom Chemstation report of the calibration standard duplicates being verified

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

Controls

1. Negative Control (Blank): Deionized Water
2. Mixed Volatiles Control: Chemicals that meet American Chemical Society (A.C.S.) specifications will be used when available. To a 500 mL volumetric flask add 500 μ L of methanol, acetone, ethanol and isopropanol and 250 μ L acetaldehyde. Dilute to volume with deionized water. The solution can be stored at room temperature and expires in 1 year.

A newly prepared mixed volatiles control is deemed suitable for use in casework when all chemicals in the mixture are detected with baseline separation during its first use and the mass spectra identify each component when compared with reference spectra. Each peak in the mixed volatiles control will be evaluated on a newly prepared standard.

After a new lot of mixed volatile control is verified the following documentation is included in the blood alcohol discipline's reagent log.

1. The chromatogram of the mixed volatile control showing baseline separation of all components
 2. The mass spectrum of each component with a library match
3. Whole Blood Ethanol Controls: Whole blood ethanol controls are purchased from a reputable supplier. A critical supply vendor form is completed for the whole blood controls. The manufacturer control sheets for all whole blood controls are kept in the blood alcohol laboratory.

Prior to using a new lot of whole blood control or a new shipment of a previously received lot, one vial of the received lot will be analyzed once (in duplicate). The averaged truncated result must fall within the manufacturer's expected range. If not, the lot will not be accepted for use. The values from the vial analyzed are entered into the Ethanol Control Worksheet. The worksheet is printed and placed with the manufacturer's control sheet and stored in the Whole Blood Controls binder. While this lot is in use the values from each consecutive use of the lot must fall within the running average +/- the expanded coverage probability of 4.2%. Values from each vial are documented in the Ethanol Control Worksheet.

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

After determining a whole blood control's parameters of precision, the following documentation is included in the Whole Blood Controls binder:

1. Certificate of Analysis
2. Control Pack (See Section 8 for requirements)
3. ETOH custom Chemstation Report for the whole blood controls being verified
4. Aqueous Ethanol Controls: NIST traceable aqueous ethanol standards at 0.025 and 0.300 g/100 mL are used as low and high level ethanol controls. Aqueous ethanol controls are purchased from an ISO/IEC 17025:2005 certified supplier. A critical supply vendor form is completed for each ethanol control. The Certificates of Analysis for all aqueous ethanol controls are kept in the blood alcohol laboratory.

Prior to using a new lot of aqueous ethanol controls, one from the lot should be run as a sample (in duplicate) to verify the lot falls within the manufacturer's stated concentration +/- the expanded coverage probability of 4.2%. See Section 3 Acceptability Requirements for further details. A new shipment of the same lot does not require additional testing. The lot numbers and expiration dates should then be entered into the LIMS where they will become part of each case file as a permanent record of the standards used for each calibration. The values from the controls analyzed are entered into the Ethanol Control Worksheet.

After a new lot of aqueous ethanol controls is verified, the following documentation is included in the Aqueous Ethanol Control binder:

1. Certificate of Analysis
2. Control Pack (See Section 8 for requirements)
3. ETOH custom Chemstation report of the aqueous ethanol standard duplicates being verified
5. Beverage Controls: Purchase a non-carbonated wine and dispense aliquots into approximately 2mL vials. Label the vials with the beverage control lot and date. The lot number is incremental. If the previous beverage control lot was lot 1, the next will be lot 2. Note the name and type of wine, the percent alcohol from the bottle's label, the transfer date and analyst that created the lot along with the lot number in the Whole Blood Control Binder.

Run the beverage control as a sample, in duplicate, in two beverage runs. The average of the two values, calculated to alcohol by volume, is used as the starting number in the beverage control worksheet. The alcohol by volume detected should be between 5-20%.

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

Each value from the two runs must fall within 4.3% of the average. The alcohol by volume values as determined are entered into the Ethanol Control Worksheet. The worksheet is printed and placed with the lot number sheet and stored in the Whole Blood Controls binder. While this lot is in use the values from each consecutive use of the lot must fall within the running average +/- the expanded coverage probability of 4.3%. Values from each vial are documented in the Ethanol Control Worksheet.

Date Documentation of Controls and Calibration Standards

When a new shipment of controls or calibration standards is received by the laboratory, the date received is marked on the certificate of analysis. After the new shipment is verified, the date in which they are put in use is marked on the certificate as well. Finally, when the last control or calibration standard of a shipment is consumed, the end of use date is marked on the certificate of analysis.

ARCHIVED 7/19/2015

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

SECTION 5: EQUIPMENT AND INSTRUMENTATION

Diluter/Dispenser

A Hamilton Microlab diluter/dispenser is used for sample dilution. The ETHANOL dilution method on the diluter/dispenser is set for a 100 µL sample with 1000 µL of internal standard diluent along with a 1000 µL wash between samples. The B100 beverage dilution method on the diluter/dispenser is set for a 50 µL sample with 4950 µL of internal standard diluent along with a 5000 µL wash between samples. The complete parameters for the methods can be found in the maintenance and calibration binders for each of the diluter/dispensers.

Each diluter/dispenser is calibrated annually by an external calibration provider that meets the requirements of accredited to ISO/IEC 17025:2005 with a scope of accreditation covering the calibration performed. Documentation of this maintenance and calibration information is kept in binders in the blood alcohol laboratory. Other maintenance and repair is performed, as needed, according to the manufacturer's recommendation and is documented in the maintenance log.

Internal performance checks are conducted on the diluter/dispensers between 5 and 6 months after the external calibration was performed. The Diluter Internal Check Worksheet that documents this internal performance check is completed, printed and stored in each diluter binder. The Diluter Internal Check Working Instructions detail the procedure to follow.

If an internal performance check does not pass, the diluter/dispenser will be removed from service until repair and external calibration is performed.

Headspace Gas Chromatograph/Mass Spectrometer

An Agilent gas chromatograph equipped with a headspace autosampler, a flame ionization detector and mass spectrometer is used for sample analysis. The instrument is interfaced with a computer and uses the Chemstation software to calibrate and analyze unknown case samples in the programmed sequence. A printed copy of the current instrument method used for analysis, along with any archived methods, is kept in the binder next to the instrument.

Repair and maintenance of the instrument and headspace autosampler is performed as needed in accordance with the manufacturer's recommendations. This is recorded in a binder kept in the instrument room.

Electronic copies of manufacturers' manuals for the diluter/dispensers and headspace gas chromatograph/mass spectrometers are located in the laboratory's internal network drive/discipline shares/chemistry/instrument manuals.

SECTION 6: ANALYSIS PROCEDURE

Sample Selection

In the alcohol testing section it is common to receive items that contain multiple containers (units) submitted as a single item. In these instances the analyst must determine which and how many of these units must be sampled and analyzed. The sampling plan listed below describes how the alcohol testing section makes sampling decisions.

Blood samples are usually collected in forensic alcohol testing kits containing two gray top tubes. Blood collected from an individual sequentially into the same type of tube can be treated as one item even when only one tube is sampled. The analyst report will reflect the number of tubes contained in the item along with the result of that item.

Blood collected into different types of tubes will be considered as separate items. The analyst should use their expertise in selecting the best tube for the analysis required and his/her report should indicate only the tube being tested. Any additional tubes in the item can be indicated in the notes field of the LIMS.

Beverage and homebrew samples can often have multiple units collected and submitted as the same item. In many instances the legal needs for the case may only require that one of these units be analyzed. In this case the analyst's report will indicate what was present in each item and what was tested. For example, two 10 mL gray top tubes containing a yellow cloudy liquid were submitted. The analyst would select one tube for analysis and their report would read 2 10 mL gray top tubes of yellow, cloudy liquid one tube analyzed. The ethanol result would then be reported for the tube that was analyzed.

In instances where more than one of the units must be sampled to meet critical volumes listed in AS 04.16.200 the analyst must perform full testing on all units required or use a statistical sampling plan. The statistical sampling plan used by the alcohol testing section is the hypergeometric method. The confidence level associated with this sampling plan must be 95% confidence that at least 90% of the units contain the analyte. Appendix C is the sampling plan employed. The decision about whether full testing is required or a sampling plan is employed will be made on a case by case basis. When a sampling plan is employed details will be provided in the notes of the case record and referenced on the report.

Sample Preparation

Remove calibration standards, control(s), and bloods (or beverages) to be analyzed from the refrigerator and allow to warm to room temperature prior to sampling. Prepare two vials for each sample and control that is to be analyzed. Set up one vial each for the blanks, mixed volatiles control and calibration standards. A blank is run as the first sample in each sequence

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

followed by the calibration standards. A blank and the mixed volatiles control is run following the calibration standards. A 0.300 aqueous ethanol control, whole blood or ethanol control and a 0.025 aqueous ethanol control are run in duplicate immediately prior to and following a batch of case samples. The maximum number of case samples per run is 15. A sample blood alcohol sequence can be found in the instrument binder located in the instrument room.

Label the autosampler vials, using indelible marker, with the appropriate lab number, calibration level, or control.

Begin sample preparation by priming the diluter/dispenser with the internal standard diluent. Ensure there is sufficient internal standard diluent in the bottle to dilute the entire run. Make sure that there are no bubbles in either syringe. Select the ETHANOL program and follow the instructions as it directs you through the setup. Ensure that each sample is mixed by inverting gently several times prior to sampling. Dispense 1000 μ L of internal standard diluent, along with 100 μ L of sample, calibration standard, or control into the headspace vials using the diluter/dispenser. Wipe the tip of the dispenser between each sample with a Kimwipe. Cover the vials with the caps and crimp tightly onto the vials. Continue this process for all samples in the run. If the quality of a sample won't allow for analysis the analyst will notify the officer that analysis was not performed.

Beverage and homebrew samples are run in separate batches than whole blood and serum samples. In a beverage/homebrew run the whole blood controls that are run before and after case samples are replaced with the beverage control. One vial of the beverage control is diluted 1:100, twice to make two primary dilutions. These dilutions are tested in duplicate, once at the beginning of the run and once at the end. The lot number of the beverage control will be listed on the Ethanol Control Worksheet that will be scanned with each case file into the LIMS.

Beverage and homebrew samples that contain ethanol may need a preliminary dilution before the diluter/dispenser process. These samples may be diluted 1:100 with deionized water. The dilution factor can then be adjusted by the analyst if needed based on the test results. The resulting ethanol concentration will then need to be multiplied by the dilution used, and then converted to volume of ethanol per volume of liquid by dividing by the density of ethanol. For example:

$$0.050 \text{ g/100 mL (ETOH concentration)} \times 100 \text{ (dilution factor)} \div 0.789 \text{ g/mL (density ETOH)} = 6.3 \text{ mL/100 mL (final ETOH concentration)}$$

In this example, the beverage would be 6.3% ethanol by volume. The "proof" would be twice that, or 12.6.

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

Instrument Preparation

Sequences are imported to the Chemstation software using the LIMS. Instructions on sequence importation and the method parameters are located in the binder next to the instrument. Place all vials into the auto sampler tray in the order they are listed in the sequence table.

Each day, prior to running a sequence, the mass spectrometer is calibrated using the autotune program. A successful calibration is indicated when the instrument is able to assign the proper masses to the Perfluorotributylamine (PFTBA) calibration standard. The tune evaluation must pass in order to use the instrument.

A copy of the Autotune report and tune evaluation is included in the control packet associated with each case in the run.

Software Usage

Sequences are imported into Chemstation using the LIMS. Instructions on sequence importation are located in a binder next to the instrument. For additional instructions on how to use the Chemstation software refer to the manufacturer's instructions.

ARCHIVED 7/19/2015

SECTION 7: REPORTING

Upon completion of the run verify that negative control or blanks demonstrate no measurable ethanol. Ensure the mixed volatiles control has baseline separation of all components and that the mass spectrum of ethanol contains the ions 31, 45, and 46. Calibration standards should read within +/- 0.005 or 5%, whichever is greater, of their expected value. The aqueous ethanol results must be within +/- 4.2% of the manufacturer's expected concentration. Whole blood control results must within +/- 4.2% of the lot's running average. Beverage control results, calculated to alcohol by volume, must fall within +/- 4.3% of the lot's running average. If a blank, mixed volatiles control, calibration standard, aqueous ethanol control, beverage control or whole blood control is out of range or does not meet the above requirements the case samples should be rediluted and rerun.

The two analyzed values for each unknown sample and all controls must agree within +/- 3%. The reported value is the average of the two analyzed values truncated to three decimal places.

The expanded combined standard uncertainty at 95% coverage for blood ethanol analysis is +/- 4.2% and +/- 4.3% for beverage ethanol analysis at the Alaska Scientific Crime Detection Laboratory. This is reported in the same units as the ethanol result in the format:

Blood Alcohol Result 0.xxx g/100mL +/- 0.xxx g/100mL

For additional information regarding measurement of uncertainty see Appendix A.

If duplicate analyzed values do not agree within +/- 3% and the mean of the duplicates is greater than the limit of reporting, then the sample will be diluted a second time and analyzed. If the second dilution's analyzed values does not fall within +/- 3%, the sample may be diluted a third time. Should the analyzed values of the third dilution not fall within +/- 3%; the sample will be reported as "Sample quality insufficient for quantitation".

"None Detected" will be reported for analyzed values less than or equal to 0.009 g/100mL. Linearity is established from 0.020 to 0.500 g/100mL. Alcohol values greater than or equal to 0.010 and less than 0.020 g/100mL will be reported as less than 0.020 g/100mL. Samples with alcohol values over 0.500 g/100mL will be diluted and rerun.

After performing the beverage alcohol calculation the concentration will be reported as follows:

- If greater than or equal to 0.025% v/v (limit of quantitation) report ethanol concentration.
- If less than 0.025% v/v but greater than or equal to 0.012% v/v report "Positive ethanol, less than 0.025% v/v".
- If less than 0.012% v/v (limit of reporting) report "None Detected".

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

The ethanol retention time is determined by the method. The sequence is designed to set the retention time of ethanol to that of the first ethanol calibrator. Subsequent peaks must fall within plus or minus 5% of that retention time per the method. Any peaks not falling in this range are not considered to be ethanol and will not be quantitated as such.

The mass spectrum generated for the ethanol peak is included on the ETOH custom Chemstation report. The analyst should ensure that the mass spectrum of ethanol contains the ions 31, 45 and 46.

A LIMS system is used for compiling all case and run data into each case file and generating crime laboratory reports. There are three types of requests in the LIMS for forensic alcohol testing: blood alcohol, beverage alcohol and serum/plasma alcohol. The analyst must edit the request in the LIMS to reflect the sample they are testing. Instructions for importing data are located in a binder next to the instrument. After the data is imported the analyst should review the imported data and case file. When the analyst determines the report is complete and ready for review he/she should record the date in the date of completion field of the notes function in the LIMS. The analyst should then affix their electronic signature.

After reports are signed by the analyst they are forwarded to a qualified person for technical and administrative review.

At the completion of analysis the glass headspace vials should be disposed of in the appropriate biohazard container and the evidence returned to the evidence section if no other testing is required.

ARCHIVED 7/19/2015

SECTION 8: CASE FILES

Case files are stored electronically in the LIMS system. Each case file is composed of the control pack which includes:

- The Autotune report and Tune Evaluation
- Calibration curve
- The ETOH custom Chemstation report for the blanks, mixed volatiles control, each calibration standard, each aqueous ethanol control and whole blood or ethanol control
- A copy of the sequence table for the run
- Ethanol Control Worksheet

The case file also includes the following in addition to the control pack listed above:

- The ETOH custom Chemstation report for each duplicate
- The analyst's worksheet
- The Laboratory's Request for Laboratory Services
- Chain of Custody Documentation
- The Crime Laboratory report
- Any case related correspondence

ARCHIVED 7/19/2015

SECTION 9: PROFICIENCY TESTING

The guideline for satisfactory completion is based on the manufacturer's expected results and the average of all respondents. This means $\pm 10\%$ or ± 2 standard deviations of the average of all respondents for samples. Proficiency test results are maintained by the Quality Assurance Manager.

Blood alcohol results provided to the proficiency test provider will be reported in the same manner as calculated for casework.

ARCHIVED 7/19/2015

SECTION 10: Providing Opinions for Court Proceedings

Providing Opinions For Court Proceedings

In general, calculations based on a blood alcohol concentration are elicited as expert opinion during testimony in court proceedings. The Interpretation of Alcohol Results manual outlines the interpretation of the results of blood and breath alcohol measurements. In some instances a written opinion is requested prior to court proceedings. In this case the analyst should prepare a memo containing their opinion and issue it to the requesting agency. For cases in which a blood alcohol analysis was performed within the laboratory, the prepared memo will be scanned into the imaging module for the relevant case. Memos not relevant to a specific analysis performed within the laboratory will be scanned into the imaging module for the case B-OPINIONYYYY (where YYYY is the year that the memo was written) located within the breath alcohol database. A report should not be generated for expert opinion.

ARCHIVED 7/19/2015

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

SECTION 11: TOXICOLOGY SEND OUTS

Toxicology services provided by Washington State Patrol to the Alaska Scientific Crime Detection Laboratory are funded through a National Highway Traffic Safety Administration (NHTSA) grant. The Alaska Scientific Crime Detection Laboratory does not currently send evidence to Washington in non-driving offenses for drug toxicology analysis. Evidence will be sent to Washington State Patrol for drug toxicology analysis if the case meets one of the following criteria:

- A Drug Recognition Expert (DRE) evaluation was performed
- The submitting agency requests a drug toxicology analysis and the case involves a driving related accident leading to the serious injury or fatality of a person other than the driver
- The submitting agency requests a drug toxicology analysis, the case involves a driving related offense, and a blood alcohol request is made
- The submitting agency requests a drug toxicology analysis, the case involves a driving related offense, a blood alcohol request is made and its result is less than 0.100 g/100 mL

Note: At the blood alcohol discipline supervisor's discretion, evidence from traffic related cases not meeting the criteria outlined above may also be sent to Washington State Patrol for drug toxicology analysis.

Before staff members prepare evidence for drug toxicology send outs, they will ensure that any blood alcohol requests for the associated cases are complete. Instructions for performing Washington toxicology send outs and entering Washington toxicology analysis reports in the LIMS are located in the blood alcohol laboratory.

SECTION 12: ABBREVIATIONS

AEC: Aqueous Ethanol Control

FID: Flame Ionization Detector

GC: Gas Chromatograph

HSS: Headspace sampler

PFTBA: Perfluorotributylamine

MS: Mass Spectrometer

MV: Mixed volatiles

RLS: Request for laboratory services form

RPD: Relative percent difference

WBC: Whole blood control

ARCHIVED 7/19/2015

APPENDIX A: UNCERTAINTY OF MEASUREMENT

The estimation of the uncertainty of measurement of the ethanol concentration in samples has been performed. The initial study performed in 2012-2013 has been documented in a report titled "Forensic Alcohol Estimate of the Uncertainty of Measurement December 20, 2013 Blood Ethanol and Beverage Ethanol". This report is stored in the internal network drive of the Forensic Alcohol discipline share folder and a copy is kept in the Quality Assurance Records.

The reports address the following elements:

- Statement of the measurand
- Statement of the measurement traceability
- Equipment used
- All uncertainty components considered
- All significant uncertainty components and their evaluation
- Data used to estimate repeatability and/or reproducibility
- All calculations
- The combined standard uncertainty, the coverage factor, the coverage probability, and the resulting expanded uncertainty
- The schedule to review and/or recalculate the measurement of uncertainty.

The current expanded combined standard uncertainty at 95% coverage for blood ethanol analysis is +/- 4.2% and +/- 4.3% for beverage ethanol analysis at the Alaska Scientific Crime Detection Laboratory.

An ongoing estimation of the uncertainty of measurement is performed utilizing data from all case batches. This information is stored in the Ethanol Control Worksheet stored in the internal network drive of the Forensic Alcohol discipline share folder.

The data from this ongoing estimation of the uncertainty of measurement will be reviewed and an updated estimation of the uncertainty of measurement reports will be written at least annually or if any significant change in the expanded uncertainty is detected.

Alaska Scientific Crime Detection Laboratory

Quantitative Alcohol Procedure Manual

Issued: 4/27/2015
Effective: 4/27/2015

Version: QAPM 2015 R0
Status: Archived

APPENDIX B: REVISION HISTORY

2015 R0	2014 R1	Location	Changes Made
all	all	where applicable	Where whole blood controls are referenced, added beverage control as well.
5	5	Section 3, Acceptability Requirements, Second paragraph	Added to the end "The beverage control value is converted to alcohol by volume, as outlined in Section 6, Sample preparation, and the value must fall within 4.3% of the expected value."
6	6	Section 4, Internal Standard Diluent	Under Internal Standard Diluent second paragraph changed "only one integrated peak" to "no ethanol is detected". Changed "to showing n-propanol and no ethanol is detected."
8-9	N/A	Section 4, Controls	Add: "5 Beverage Controls
12	11	Section 6, Sample Preparation, Following third paragraph	New paragraph added on beverage and homebrew samples.
14	13	Section 7, Reporting, Paragraph 1, Following 5th sentence	Sentence added: "Beverage control results, calculated to alcohol by volume, must fall within +/- 4.3% of the lot's running average."
14	13	Section 7, Reporting, Paragraph 3, 1st sentence	Remove "5.1%" and replace with "4.3%".
14	13	Section 7, Reporting	Second bulled of beverage alcohol calculation updated.
21	20	Appendix A, Paragraph 3	Updated the first sentence to "current" expanded combined uncertainty..." Remove "5.1%" and replace with "4.3%".

ARCHIVED 7/19/2015

APPENDIX C Hypergeometric Sampling Plan*

Based on statistical probability, there is 95% confidence that at least 90% of the units contain ethanol.

The use of the sampling plan requires that all units appear to be homogenous.

If any results are different than the rest, the analyst must re-evaluate the population.

The report will reference the sampling plan utilized and the confidence levels and corresponding inferences of the population. The report will state what was received and clearly state the results/conclusions.

Population Size	Sample Size
1-12	All
13-20	12
21-30	15
31-40	18
41-50	19
51-60	20
61-70	21
71-80	22
81-100	23
101-130	24
131-180	25
181-270	26
271-470	27
471-1000	28

**Information is taken from enfsi-dwg sampling calculator 2012 referenced in the ASCLD/LAB Policy on Sampling, Sampling Plans and Sample Selection.*

ARCHIVED 7/19/2015