

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument
State of Alaska

JUL 15 2010

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N: <u>130126</u> ✓
Name: <u>Farmer</u>	ID#: <u>6450</u> ✓	Date: <u>7/9/10</u> ✓
A Agency: <u>Anchorage Airport Police</u>		Phone #: <u>266-2411</u>
Instrument Location: <u>ANCHORAGE AIRPORT POLICE DEPT.</u>		
Alco S/N: <u>X172961</u> ✓		Target Value: <u>.079</u> ✓
		High Pressure: <u>800 psi</u>
B Alco Test Value Average: <u>.078</u> ✓		<u>.078</u> ✓
		1 st Alco 2 nd Alco
Signature: <u>[Signature]</u> <u>MAG</u>		
(OVER)		

BMB
7/15/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27th day of July, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office
BT6 9/95



