

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 11/7/2022

Instrument #: 100387

Instrument Location: Palmer PD

Supervisor Name and Agency: Sgt. Luke Szipszky (7042), Palmer PD

Describe the Status Change or Issue with the Instrument:

Displaying several error messages "undefined error" and "unable to load ticket database"

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 11/4/2022 9:00 PM

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor