

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 2/2/2015
Effective: 2/2/2015

Version: DMCIF 2015 R0
Status: Active

Instrument S/N 100351

Received From ARD

Date 4/4/15

Analyst Colleen O'Bryant

1. Reason for instrument return:

Breath hose removed at ARD while instrument was in service. Unable to determine original breath hose that instrument was certified with. Removed instrument from service

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Visual Inspection

5. Upload

6. Update agency information and polling status in DM Host Calling Program

7. Check for VOCs

From NPAS

8. NPAS Repair Paperwork (attached)

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/07/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 26.91: 0.077
 LOT #: 09214080A2
 EXPIRATION: 05/01/2016
 TANK PRESSURE: 247 psi

BLANK TEST 0.000 11:04
 INTERNAL STANDARD VERIFIED 11:04

Gas Flow Error

Diagnostic Check

VERSIONS

DMT: 3.00
 PIC: 3.03
 Modem: 2.6
 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
 Breath Tube = 47.3°C PASSED

PUMP INFO

Flow Rate = 4.821 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
 PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
 Filter 2 PASSED
 Filter 3 PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 20 _____

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/07/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 28.92: 0.077
 LOT #: 11313080A1
 EXPIRATION: 06/01/2015
 TANK PRESSURE: 928 psi

BLANK TEST	0.000	11:11
INTERNAL STANDARD	VERIFIED	11:11
EXTERNAL STANDARD	0.076	11:11
BLANK TEST	0.000	11:12
EXTERNAL STANDARD	0.077	11:12
BLANK TEST	0.000	11:13
EXTERNAL STANDARD	0.077	11:13
BLANK TEST	0.000	11:14
EXTERNAL STANDARD	0.077	11:14
BLANK TEST	0.000	11:15
EXTERNAL STANDARD	0.077	11:16
BLANK TEST	0.000	11:16

Average = 0.0768
 Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.00
 PIC: 3.03
 Modem: 2.6
 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
 Breath Tube = 45.1°C PASSED

PUMP INFO

Flow Rate = 4.750 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
 PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
 Filter 2 PASSED
 Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Boiz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 Nita J. Boiz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this ____ day of _____, 20 ____

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument**State of Alaska*

Serial #: 100351

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/11/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.00: 0.078

LOT #: 11313080A1

EXPIRATION: 06/01/2015

TANK PRESSURE: 922 psi

BLANK TEST 0.000 12:02

INTERNAL STANDARD VERIFIED 12:02

Gas Flow Error

Diagnostic Check

VERSIONS

DMT: 3.00

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED

Breath Tube = 45.8°C PASSED

PUMP INFO

Flow Rate = 4.815 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this ____ day of _____, 20 ____