

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: July 10, 2020 Instrument Serial Number: 100364

Instrument Location: Kodiak Police Department

Supervisor Name and Agency: Troopers Anthony Will and Derek Beaver; Alaska State Troopers Kodiak Post

Describe the Status Change or Issue with the Instrument:

DMT 100364 is malfunctioning and is being sent back to the crime lab for repairs or replacement.

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: July 10, 2020 @1249 hours.

For Use by SCDL

Additional Notes

---

---

---

---

---

---

---

---

---

---

*Touch screen not working jmv 08/03/20*

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740