

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100354

Received From: APD

Date: 3/25/25

Analyst: Megan Carbeck

Reason for instrument return:

Data files are exporting to DM Host incorrectly. The printable PDFs for 4 breath tests are available on the instrument, however they are not present in the raw data files. Instrument returned with the tank present.

From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐