

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100399

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/12/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 28.98: 0.077

LOT #: AG310901

EXPIRATION: 04/19/2025

TANK PRESSURE: 129 psi

BLANK TEST	0.000	12:05
INTERNAL STANDARD	VERIFIED	12:05
EXTERNAL STANDARD	0.080	12:06
BLANK TEST	0.000	12:07

Gas Flow Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C

PASSED

Breath Tube = 46.9°C

PASSED

PUMP INFO

Flow Rate = 4.565 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED

File Only DM 01/02/25