

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100676

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/30/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.66: 0.079  
LOT #: 01315080A2  
EXPIRATION: 02/05/2017  
TANK PRESSURE: 1028 psi

BLANK TEST	0.000	14:37
INTERNAL STANDARD	VERIFIED	14:37
EXTERNAL STANDARD	0.078	14:37
BLANK TEST	0.000	14:38
EXTERNAL STANDARD	0.078	14:39
BLANK TEST	0.000	14:39
EXTERNAL STANDARD	0.078	14:40
BLANK TEST	0.000	14:41
EXTERNAL STANDARD	0.078	14:41
BLANK TEST	0.000	14:42
EXTERNAL STANDARD	0.078	14:42
BLANK TEST	0.000	14:43

Average = 0.0780  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 48.1°C PASSED

### PUMP INFO

Flow Rate = 4.086 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

10/7/2015

Subscribed and sworn before me this 7<sup>th</sup> day of Oct, 20 15

  
Nikki Roth, Notary Public  
My Commission Expires With Office



CRF  
10/2/15