

	State of Alaska Scientific Crime Detection Laboratory		
	<b>DataMaster Calibration Form – DMCAF2011R1</b>		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

**DataMaster DMT Calibration Form**

Instrument # 100417

Date: 6/24/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Repro
- c. Lot 10003
- d. Expiration Date 9/8/12

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. O'Bryant

*COB 1/2*

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100417  
-----

Date: 06/24/2011  
Time: 07:29:47

SUPERVISOR NAME:  
COLLEEN S OBRYANT  
CERTIFICATE #: 6237

LOT #: REPC0100  
BOTTLE #: 10003  
EXPIRATION: 09/08/2012

Ca = 0.1000  
CAL = 0.962023    0.800 <= CAL < 1.200  
b1 = 0.0000    0.0000 <= b1 < 0.0040  
b2 = 0.0043    0.0010 <= b2 < 0.0100  
b3 = 0.0000    0.0000 <= b3 < 0.0040  
Xq = 0.0944    0.0500 <= Xq < 0.2000  
a21 = 1.145812    1.050 <= a21 < 1.300  
a31 = 0.525561    0.300 <= a31 < 0.800

*CUB 2/2*

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100417\B-100417 ECF-20140

2/18/2014