

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100667
Start Date: 11/7/23
Date Received: 11/7/23

Analyst(s): Carlie Glaister

Reason for instrument evaluation

Instrument not recognizing when a breath sample is being provided

Visual Inspection

- | | |
|---|---|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box (modem) is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Double Check Valve secured to wall | <input checked="" type="checkbox"/> All tubing and connections secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag | |

As Found Conditions:

Barometric Pressure (inHg): 29.19 Temperature (C): 22.6 Relative Humidity (%): 18

Linearity Test (attached)

0.020g/210L Mnf: Airgas	Lot: AG217902	Exp: 6/28/24
0.080g/210L Mnf: Airgas	Lot: AG310901	Exp: 4/19/25
0.300g/210L Mnf: Airgas	Lot: AG217903	Exp: 6/28/24

Notes

5 way valve was stuck. Unmoved the mechanism to un-jam it. While testing the mechanism, it continues to get stuck. Recommend replacing 5 way valve. Reset options.

- | | |
|---|--|
| <input checked="" type="checkbox"/> VOC (attached) | <input checked="" type="checkbox"/> Verify software version |
| <input checked="" type="checkbox"/> VOC Internal Standard value ≤ 3.00% | <input checked="" type="checkbox"/> Non-Drinking Subject Test (attached) |
| Condition of item: <input type="checkbox"/> Good | <input checked="" type="checkbox"/> Needs Repair
5 way valve got stuck & would not recognize breath sample resulting in an incomplete . |
| <input type="checkbox"/> Ready for Adjustment | |
| <input type="checkbox"/> Ready for Calibration | |

End Date: 11/7/23

LINEARITY CHECK


Alaska Department of Public Safety
DATAMASTER dmt: 100667

Date: 11/07/2023
Time: 12:47:02

SUPERVISOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

TARGET	AVERAGE	STD. DEV.
0.020	0.0200	0.0000
0.078	0.0772	0.0004
0.293	0.2928	0.0004

SLOPE = 1.000226
 R^2 = 0.999992



LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety
DATAMASTER dmt: 100667


Date: 11/07/2023
Time: 12:47:04

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
TARGET AT 29.19: 0.020
LOT #: AG217902
EXPIRATION: 06/28/2024
TANK PRESSURE: 733 psi

BLANK TEST	0.000	12:48
INTERNAL STANDARD	VERIFIED	12:49
EXTERNAL STANDARD	0.020	12:49
BLANK TEST	0.000	12:50
EXTERNAL STANDARD	0.020	12:50
BLANK TEST	0.000	12:51
EXTERNAL STANDARD	0.020	12:51
BLANK TEST	0.000	12:52
EXTERNAL STANDARD	0.020	12:52
BLANK TEST	0.000	12:53
EXTERNAL STANDARD	0.020	12:53
BLANK TEST	0.000	12:54

Average = 0.0200
Std Dev = 0.0000



LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety
DATAMASTER dmt: 100667






Date: 11/07/2023
Time: 12:56:18

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.19: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 713 psi

BLANK TEST	0.000	12:57
INTERNAL STANDARD	VERIFIED	12:57
EXTERNAL STANDARD	0.077	12:58
BLANK TEST	0.000	12:59
EXTERNAL STANDARD	0.078	12:59
BLANK TEST	0.000	13:00
EXTERNAL STANDARD	0.077	13:00
BLANK TEST	0.000	13:01
EXTERNAL STANDARD	0.077	13:01
BLANK TEST	0.000	13:02
EXTERNAL STANDARD	0.077	13:02
BLANK TEST	0.000	13:03

Average = 0.0772
Std Dev = 0.0004



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety
DATAMASTER dmt: 100667






Date: 11/07/2023
Time: 13:25:19

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300
TARGET AT 29.18: 0.293
LOT #: AG217903
EXPIRATION: 06/28/2024
TANK PRESSURE: 285 psi

BLANK TEST	0.000	13:26
INTERNAL STANDARD	VERIFIED	13:26
EXTERNAL STANDARD	0.293	13:27
BLANK TEST	0.000	13:28
EXTERNAL STANDARD	0.293	13:28
BLANK TEST	0.000	13:29
EXTERNAL STANDARD	0.293	13:29
BLANK TEST	0.000	13:30
EXTERNAL STANDARD	0.292	13:30
BLANK TEST	0.000	13:31
EXTERNAL STANDARD	0.293	13:31
BLANK TEST	0.000	13:32

Average = 0.2928
Std Dev = 0.0004



VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100667

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/07/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.16: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 729 psi

BLANK TEST	0.000	10:08
INTERNAL STANDARD	VERIFIED	10:08
Xq = 0.105	(0.91%)	
EXTERNAL STANDARD	0.077	10:08
X[1] = 0.0774	(-0.0025)	(-0.0017)
BLANK TEST	0.000	10:09
EXTERNAL STANDARD	0.077	10:09
X[1] = 0.0773	(-0.0025)	(-0.0012)
BLANK TEST	0.000	10:10
EXTERNAL STANDARD	0.077	10:10
X[1] = 0.0772	(-0.0026)	(-0.0020)
BLANK TEST	0.000	10:11
EXTERNAL STANDARD	0.077	10:11
X[1] = 0.0772	(-0.0021)	(-0.0013)
BLANK TEST	0.000	10:12
EXTERNAL STANDARD	0.077	10:13
X[1] = 0.0775	(-0.0026)	(-0.0010)
BLANK TEST	0.000	10:13

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8 °C	PASSED
(44.0 °C - 52.0 °C)	
Breath Tube = 45.6 °C	PASSED
(38.0 °C - 50.0 °C)	

SETTINGS

Lamp Voltage = 2.15 V	PASSED
(1.15 V - 2.67 V)	
Cooler Voltage = 1.75 V	PASSED
(1.16 V - 2.70 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 517 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.2 in	

PUMP INFO

Flow Rate = 4.693 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.003 V <= 0.010 V)	
PUMP OFF	PASSED
(0.003 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.106 0.17% (0.00% - 4.00%)	PASSED
----------------------------------	--------

40.00

36.00

32.00

28.00

24.00

20.00


16.00

12.00

8.00

4.00

0.00



SUBJECT TEST	
Alaska Department of Public Safety DATAMASTER dmt: 100667	
Date:	11/07/2023
Time:	14:01:26
SUBJECT NAME:	
TEST TEST	
LICENSE #:	XXXX
OPERATOR NAME:	CARLIE GLAISTER
OPERATOR'S #:	10436
DEPT/AGENCY:	
NON-DRINKING SUBJECT TEST	
CASE/REPORT:	XXXX
TEST TYPE:	TEST
EXTERNAL STANDARD INFORMATION	
NOMINAL:	0.080
TARGET AT 29.18:	0.078
LOT #:	AG310901
EXPIRATION:	04/19/2025
DIAGNOSTIC CHECK	PASSED
BLANK TEST	0.000
INTERNAL STANDARD	VERIFIED
EXTERNAL STANDARD	0.078
BLANK TEST	0.000
SUBJECT SAMPLE	INCOMPLETE
	14:02
	14:03
	14:03
	14:03
	14:04
	14:06
	16.00
	12.00
	8.00
	4.00
	0.00
0 10 20 30 40 50 60 70 80 90 100 110 120	