

# Alaska Scientific Crime Detection Laboratory

## DataMaster Acceptance Form

Issued: 01/27/2012  
Effective: 01/27/2012

Version: DMAF2012 R0  
Status: Active

Instrument # 100692

Date: 11/21/12

Analyst: Colleen O'Bryant

- Check in instrument in to JusticeTrax Breath Alcohol Database
- Visual Inspection
- Update software  (2.00)
- Print calibration factors (attached)
- NPAS calibration paperwork (attached)
- Barcode Instrument
- State of Alaska Property ID# 12-F0383

Breath



QSWNV3NEOVKX2G  
B-100692



*Certificate of Calibration*

100692

Instrument Serial Number

10260

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

11-3-2011

Date

A handwritten signature in black ink, appearing to be "S. J.", written over a horizontal line.

Technician

## CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100692  
-----

Date: 11/02/2011  
Time: 08:49:23

|     |   |          |        |                |
|-----|---|----------|--------|----------------|
| Ca  | = | 0.1000   |        |                |
| CAL | = | 0.993327 | 0.800  | <= CAL < 1.200 |
| b1  | = | 0.0000   | 0.0000 | <= b1 < 0.0040 |
| b2  | = | 0.0036   | 0.0010 | <= b2 < 0.0100 |
| b3  | = | 0.0000   | 0.0000 | <= b3 < 0.0040 |
| Xq  | = | 0.0906   | 0.0500 | <= Xq < 0.2000 |
| a21 | = | 1.191929 | 1.050  | <= a21 < 1.300 |
| a31 | = | 0.590430 | 0.300  | <= a31 < 0.800 |

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100692\B-100692 ECF-20140

7/10/2014