

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100255

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 05/20/2025

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 2.05  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 42.2°C	PASSED

Detector Overflow

File Only dmt 06/18/25