

Alaska Scientific Crime Detection Laboratory

DataMaster Repair Form

Effective: 3/20/2025

Version: 2.0

Instrument #: 100355

Analyst(s): Kaya Swisher

Start Date: 3/21/25

Reason for Repair

Internal standard cleaning-

Troubleshooting/Repair

Cleaned internal standard with IPA on 3/21/25
and again on 6/2/25. Ready for adjustment, VOC
and Non-Drinking test not needed at this time.

☐ VOC (attached)

☐ VOC Internal Standard value $\leq 3.00\%$

☐ Verify software version

☐ Non-Drinking Subject Test (attached)

☒ Ready for Adjustment

☐ Further Repair Needed

☐ Proceed to Close Out

☐ Ready for Calibration

☐ Return to Factory

KS 6/16/25

End Date: 6/16/25