

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument  
State of Alaska*

Serial #: 100675

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 03/13/2025

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 48.1°C PASSED

### PUMP INFO

Flow Rate = 4.932 L/M PASSED

### DETECTOR INFO

PUMP ON FAILED  
PUMP OFF FAILED

### FILTER INFO

Filter 1 FAILED  
Filter 2 FAILED  
Filter 3 PASSED

### INTERNAL STANDARD

PASSED

Diagnostic Check Failed

File only  
CNF 5/12/25