

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100374

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/18/2013

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.060
TARGET AT 29.78: 0.079
LOT #: 02012080A1
EXPIRATION: 03/01/2014

BLANK TEST	0.000	23:06
INTERNAL STANDARD	VERIFIED	23:06
EXTERNAL STANDARD	0.078	23:06
BLANK TEST	0.000	23:07
EXTERNAL STANDARD	0.078	23:07
BLANK TEST	0.000	23:08
EXTERNAL STANDARD	0.078	23:09
BLANK TEST	0.000	23:09
EXTERNAL STANDARD	0.078	23:10
BLANK TEST	0.000	23:11
EXTERNAL STANDARD	0.078	23:11
BLANK TEST	0.000	23:12

Average = 0.0780
Std Dev = 0.0000

Diagnostic Check

VERSIONS
DMT: 2.00
PIC: 3.02
Modem: 2.3
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 5.331 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

8/13/2013

Subscribed and sworn before me this 13 day of Aug 20 13


Nikki Roth, Notary Public
My Commission Expires With Office



CRB 8/9/13