

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 1/18/2024

Version: 4.0

Instrument #: 100390

Analyst/Technician(s): Kayla Carlson

Start Date: 3/23/26

Date Instrument Received: 3/23/26

Reason for instrument evaluation

Filter wheel errors.

Visual Inspection:

Clean

Replace missing screws

Cermetex Box (modem) is secure

Ensure modem cable present

Double Check Valve secured to wall

All tubing and connections secure

O-ring inspection and ensure 2 replacement O-rings in bag

Attach VOC VOC Internal Standard $\leq 3.00\%$

Verify software version

Condition of item: No apparent damage Damaged

NOTE: As Found Linearity is required if the instrument is not damaged and the internal standard value is not more than 4.00%

As Found Conditions (if performed):

Barometric Pressure (inHg): _____ Temperature (C): _____ Relative Humidity (%) _____

Linearity Test Levels (g/210L): 0.020, 0.080, 0.300

Notes: VOC & Linearity not completed due to filter wheel errors.

Attach Non-Drinking Subject Test KC 3/23/26

Conclusion:

Ready for Adjustment

Needs Repair

Ready for Calibration

Lab Work End Date: 3/23/26

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100390

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/23/2026

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

Filter Wheel Error

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100390

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/23/2026

External Standard Test Values

Filter Wheel Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.6°C

PASSED

Breath Tube = 44.1°C

PASSED

PUMP INFO

Flow Rate = 4.958 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED