

Alaska Scientific Crime Detection Laboratory

DataMaster Close Out Form

Effective: 3/20/2025

Version: 3.0

Instrument #: 100396

Date: 2/6/2026

Analyst: Kaja Swisher

Reason for Instrument Close Out

Filter wheel errors, at lab since 2022. Repairs unsuccessful.

State of Alaska Property ID#: 12-F0352

Change instrument location to Close Out in JusticeTrax Breath Alcohol Database

