

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100414

Received From: APD

Date: 7/1/25

Analyst: Carlie Glaister

Reason for instrument return:

Not registering breath flow

### From the field

Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database ☒

No case CWG 7/1/25

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

### From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐