

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 4/30/2025

Instrument #: 100697

Instrument Location: Anchorage Jail

Supervisor Name and Agency: Jason Cusack, Anchorage Police

Describe the Status Change or Issue with the Instrument:

On 4/29/25, I was contacted by the crime lab and asked to remove instrument 100697 from service due to internal standard out of range status messages received recently. I conducted a VOC and then removed the instrument from service at 0620 hours on 4/30/25.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 4/30/2025 0620

-----  
**For Use by SCDL**

**Additional Notes**

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor