

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 6-5-18 Instrument Serial Number: 100685

Instrument Location: TOK AST

Supervisor Name and Agency: Sergeant Christopher Fitz / TOK AST

Describe the Status Change or Issue with the Instrument:

Out of Service Due to Issues. Crime Lab is aware

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 6-5-18 @ 1301 hrs.

For Use by SCDL

Additional Notes

Communication errors, pump errors. Instrument replaced.
JM 06/05/18

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740