

Alaska Scientific Crime Detection Laboratory

DataMaster Close Out Form

Effective: 3/20/2025

Version: 3.0

Instrument #: 100359

Date: 5/23/25

Analyst: Carlie Glaister

Reason for Instrument Close Out

Instrument was repaired by Intox for filter wheel errors. Instrument continued to have filter wheel errors shortly after being returned back to the crime lab from Intox. Instrument being closed out due to unsuccessful attempts to remedy filter wheel errors.

State of Alaska Property ID#: 10140669

Change instrument location to Close Out in JusticeTrax Breath Alcohol Database



DIAGNOSTIC RECORD

Alaska Department of Public Safety
DATAMASTER dmt: 100359

Date: 06/06/2024
Time: 09:07:44

VERSIONS
DMT: 3.02
PIC: 3.07
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.6°C	PASSED
Breath Tube = 43.0°C	PASSED

PUMP INFO	
Flow Rate = 4.411 L/M	PASSED

DETECTOR INFO	
PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO	
Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

Filter Wheel Error

DIAGNOSTIC RECORD

Alaska Department of Public Safety
DATAMASTER dmt: 100359

Date: 06/06/2024
Time: 09:10:59

VERSIONS
DMT: 3.02
PIC: 3.07
Modem: 2.6
Questions: 2.2

Filter Wheel Error