

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 08/04/2015 Instrument Serial Number: 100406

Instrument Location: Wildwood Pretrial Facility

Supervisor Name and Agency: Officer Todd Hamilton; Kenai Police Department

Describe the Status Change or Issue with the Instrument:

The battery of the UPS for the Datamaster died and needed replacement. The replacement did not work and another one had to be installed before the instrument could be put back into service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 08/02/2015 2130 hours

Out of Service Date/Time: 07/11/2015 2200 hours

**For Use by SCDL**

**Additional Notes**

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740