

Alaska Scientific Crime Detection Laboratory

DataMaster Close Out Form

Effective: 3/20/2025

Version: 3.0

Instrument #: 100387

Date: 5/23/25

Analyst: Carlie Glaister

Reason for Instrument Close Out

Instrument was previously repaired by Intox for filter wheel errors. Instrument later demonstrated internal standard & filter wheel errors that was not able to be readily remedied. Numerous attempts to address filter wheel errors by Intox & the crime lab have not been successful.

State of Alaska Property ID#: 10140882

Change instrument location to Close Out in JusticeTrax Breath Alcohol Database



VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100387

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/05/2024

External Standard Test Values

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C

PASSED

Breath Tube = 46.0°C

PASSED

PUMP INFO

Flow Rate = 4.371 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

Filter Wheel Error