

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 12/31/2024

Instrument #: 100675

Instrument Location: AST/Angoon

Supervisor Name and Agency: Luke Lemieux

Describe the Status Change or Issue with the Instrument:

See below. KS 12/31/2024

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 12/12/2024

-----  
**For Use by SCDL**

Additional Notes

The Angoon office heating went out. The VOC could not be completed for 12/12/24 due to breath hose temperature being too low. Until the office heating is back up the VOC can not be run. See emails attached. KS 12/31/2024

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

## Swisher, Kaija L (DPS)

---

**From:** Swisher, Kaija L (DPS)  
**Sent:** Tuesday, December 31, 2024 1:05 PM  
**To:** LeMieux, Luke A (DPS)  
**Cc:** Catacutan, Paulyne C (DPS)  
**Subject:** RE: VOC needed

Good afternoon Luke,

I will be filling out a Change in Status Form for this instrument documenting what is below. You don't need to fill anything out.

Just keep one of us updated when the room is back to normal temp and you have time to run a VOC again.

Happy New Year!

Kaija Swisher  
Forensic Scientist II – Chemistry Section  
Alaska Scientific Crime Detection Laboratory  
4805 Dr. MLK Jr. Ave., Anchorage, AK 99507  
Work: 907-269-5686

---

**From:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Sent:** Monday, December 23, 2024 2:57 PM  
**To:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Subject:** Re: VOC needed

Ok sounds good. I'll try it again when I go out.

Get [Outlook for iOS](#)

---

**From:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Sent:** Monday, December 23, 2024 1:17:18 PM  
**To:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Subject:** RE: VOC needed

Unfortunately, the DataMaster won't work until the room has warmed up.



**Paulyne C Catacutan**

Direct: +1 907-269-3746

Forensic Technician

[State of Alaska Scientific Crime Detection Laboratory](#)

**NOTE:** This message is intended exclusively for the individual or entity to which it is addressed. This communication may contain information that is law enforcement sensitive, proprietary, privileged, confidential and may be legally protected or otherwise exempt from disclosure. If you have received this message in error, please notify the sender immediately by email and delete all copies of this message.

---

**From:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Sent:** Monday, December 23, 2024 11:09 AM  
**To:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Subject:** Re: VOC needed

The failed test keeps happening on the detector info. Both times they failed.

Get [Outlook for iOS](#)

---

**From:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Sent:** Monday, December 23, 2024 10:50:41 AM  
**To:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Subject:** RE: VOC needed

The breath hose temp is probably from the room being too cold. Sounds good, thank you so much for helping out with this.

Happy Holidays!



**Paulyne C Catacutan**

Direct: +1 907-269-3746

Forensic Technician

[State of Alaska Scientific Crime Detection Laboratory](#)

**NOTE:** This message is intended exclusively for the individual or entity to which it is addressed. This communication may contain information that is law enforcement sensitive, proprietary, privileged, confidential and may be legally protected or otherwise exempt from disclosure. If you have received this message in error, please notify the sender immediately by email and delete all copies of this message.

---

**From:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Sent:** Monday, December 23, 2024 10:47 AM  
**To:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Subject:** Re: VOC needed

Office Toyo stove is out of fuel and probably too cold in the office. Will have fuel delivered hopefully tomorrow. I will make another trip out and run the VOC.

Get [Outlook for iOS](#)

---

**From:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Sent:** Monday, December 23, 2024 10:34:14 AM  
**To:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Subject:** Re: VOC needed

Diagnostic check failed. Then when I try to run another it says breath hose temp is too low.

Get [Outlook for iOS](#)

---

**From:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Sent:** Monday, December 23, 2024 8:02:43 AM  
**To:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Subject:** RE: VOC needed

Good morning,

We need the one from Angoon.

Thank you!



**Paulyne C Catacutan**

Direct: +1 907-269-3746

Forensic Technician

[State of Alaska Scientific Crime Detection Laboratory](#)

***NOTE:** This message is intended exclusively for the individual or entity to which it is addressed. This communication may contain information that is law enforcement sensitive, proprietary, privileged, confidential and may be legally protected or otherwise exempt from disclosure. If you have received this message in error, please notify the sender immediately by email and delete all copies of this message.*

---

**From:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Sent:** Sunday, December 22, 2024 10:02 AM  
**To:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Subject:** RE: VOC needed

Can you let me know for which Datamaster either Angoon or Kake?

---

**From:** Catacutan, Paulyne C (DPS) <[paulyne.catacutan@alaska.gov](mailto:paulyne.catacutan@alaska.gov)>  
**Sent:** Thursday, December 19, 2024 1:58 PM  
**To:** LeMieux, Luke A (DPS) <[luke.lemieux@alaska.gov](mailto:luke.lemieux@alaska.gov)>  
**Subject:** VOC needed

Good afternoon,

VOC occurred on 12/12 and your VOC is still needed. Please send that as soon as possible.

Thank you!



**Paulyne C Catacutan**

Direct: +1 907-269-3746

Forensic Technician

[State of Alaska Scientific Crime Detection Laboratory](#)

***NOTE:** This message is intended exclusively for the individual or entity to which it is addressed. This communication may contain information that is law enforcement sensitive, proprietary, privileged, confidential and may be legally protected or otherwise exempt from disclosure. If you have received this message in error, please notify the sender immediately by email and delete all copies of this message.*