

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100400

Received From: Bethel PD

Date: 1/3/2025

Analyst: Carlie Glaister

Reason for instrument return:

Tank Interference

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐