Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Version: CIISF 2014 R0

Effective: 6/11/2014 Status: Active Instrument Serial Number: 100667 Instrument Location: N Supervisor Name and Agency: Describe the Status Change or Issue with the Instrument: -enaud DD697 If instrument was placed in service or removed from service, fill out the appropriate sections below. In Service Date/Time: 4/5/11e 2200 Out of Service Date/Time: For Use by SCDL Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Issued: 6/11/2014

Foster, Charles R (DPS)

From:

DPS SCDL Toxicology (DPS sponsored)

Sent:

Wednesday, April 06, 2016 8:31 AM

To:

Foster, Charles R (DPS); O'Bryant, Colleen S (DPS); Barnett, Brandi M (DPS)

Subject:

FW: MSW DataMaster

From: Cook, Dugger A (DPS)

Sent: Tuesday, April 05, 2016 10:56 PM **To:** DPS SCDL Toxicology (DPS sponsored)

Subject: MSW DataMaster

The removed DataMaster from MSW is going to be at Palmer AST and is ready to be taken back to the crime lab. The replacement read interface detected during the first VOC but tested fine for the second. I also ran a test on myself and it proved to work fine with no additional status messages. This may be an RFI antenna problem, but I did not trouble shoot that, I am simply guessing... Thanks for getting the replacement to us so quickly.

As for the returning D/M, I can ask around the office for a delivery boy, or I can bring it in on Thursday while I am at HQ. If you need it prior to Thursday let me know, if I don't hear from you I will haul it in myself then.

Prooper D. Cook

Alaska State Troopers "B" Detachment, Palmer, Patrol 453 S. Valley Way Palmer, AK 99645 (907) 745-2131 (office) (907) 746-9164 (direct) (907) 775-0593 (cell)

> 100667 CNF 4/6/16