

Alaska Scientific Crime Detection Laboratory

DataMaster Close Out Form

Effective: 3/20/2025

Version: 3.0

Instrument #: 1003 S3

Date: 2/6/26

Analyst: Kaja Swisher

Reason for Instrument Close Out

Filter wheel errors, repair at intx unsuccessful.

State of Alaska Property ID#: 10140900

Change instrument location to Close Out in JusticeTrax Breath Alcohol Database

