

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N ¹⁰⁰³⁶⁷
~~100689~~ _{08/15/15}

Received From Kotzebue PD

Date 12/7/15

Analyst Colleen Bryant

1. Reason for instrument return:

Interference on subject test, and standard out of range.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload instrument wasn't upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From NPAS

7. NPAS Repair Paperwork (attached)